## REASONS FOR ALLOWANCE

1. The following is an examiner's statement of reasons for allowance: the claims define over the prior art of record because the prior art of record fail to teach or make obvious a method for detecting an analyte by conducting a step-wise reaction on a first portion of a sample and a simultaneous reaction on a second portion of the same sample using the same set of antibody. The ratio of the first and second reactions is used to determine the amount of analyte in the sample.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance"

Any inquiry concerning this communication or earlier communications from the
examiner should be directed to Bao-Thuy L. Nguyen whose telephone number is (571)
 272-0824. The examiner can normally be reached on Monday -- Thursday from 9:00
 a.m. - 3:00 p.m.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Long V. Le can be reached on (571) 272-0823. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 1641

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/Bao-Thuy L. Nguyen/ Primary Examiner, Art Unit 1641 09 April 2008